

PBS-Waschlösung / Washing Solution ELISA

REF: 141201

WASHBUF 20x

Instruction For Use

FOR IN VITRO DIAGNOSTICS ONLY
For professional use only

Virotech Diagnostics GmbH
Waldstrasse 23 A2
63128 Dietzenbach, Germany

Tel.: +49(0)6074-23698-0
Fax.: +49(0)6074-23698-900
www.goldstandarddiagnostics.com



1. INTENDED PURPOSE

WASHBUF 20x	
Detection/Measurement	For use with ELISA products from Virotech Diagnostic
Function	Reagent for washing the MTP between the individual reaction steps
Classification	Class A
Automation	n/a
Type	n/a
Specimen	n/a
Testing population	n/a
Intended user	Specialised personnel in laboratories

2. TEST PRINCIPLE

WASHBUF 20x is used to wash the microplate between the different steps of the procedure following the instruction for use of the assay.

3. PACKAGE CONTENTS

WASHBUF 20x 50ml, pH 7,2, with preservative and Tween 20, must be diluted before use

Quality Control Certificate

Hazard information

The Safety Data Sheet (SDS) is available at: www.virotechdiagnostics.com

4. STORAGE AND STABILITY OF THE READY-TO-USE COMPONENT

Store **WASHBUF 20x** at 2-8°C. The shelf life of the reagent is stated on the respective label. After opening, the reagent is stable for three months. After dilution, the product is stable for 4 weeks.

5. PREPARATION OF THE REAGENT

Before use, bring the reagent to room temperature (18-25°C) and mix well by inverting several times. **WASHBUF 20x** must be diluted before use.

6. SAMPLE DILUTION, PROCEDURE, INTERPRETATION OF RESULTS AND PERFORMANCE CHARACTERISTICS

The reagent is diluted and used according to the instruction for use of the ELISA products of Virotech Diagnostics before starting the test.

Attention: To avoid microbial contamination, the following points must be observed:

- Use only clean pipette tips, dispensers, and laboratory materials
- Do not return residual liquid to original bottle
- Do not change the lid
- Immediately reseal bottles after opening and removal
- Check reagents for contamination when used again

The further performance and evaluation of the test is also carried out according to the information in the instructions for use of the ELISA product. The respective performance characteristics of the product can also be found here.

7. QUALITY CONTROL

In accordance with Virotech Diagnostics GmbH's ISO-certified quality management system, each batch of **WASHBUF 20x** is tested against predetermined specifications to ensure consistent product quality.

8. LIMIT

WASHBUF 20x must not be mixed with or substituted by reagents of other manufacturers.

9. WARNINGS AND PRECAUTIONS





- In vitro diagnostic agent for professional use.
- The information, safety measures and warnings in the instructions for use must be strictly followed. In case of deviation, the user is liable for any incorrect results.
- Wearing a lab coat, disposable gloves and protective goggles is recommended during use. If contact with reagent nevertheless occurs, the relevant Safety Data Sheet (SDS) should be consulted.
- All materials of human or animal origin are to be considered potentially infectious and treated accordingly.
- Do not use after expiry date.

10. DISPOSAL CONSIDERATIONS









Information on the Reagents

Chemicals and preparations, as well as their containers, are generally hazardous waste. Their disposal is subject to national waste legislation and regulations. The competent authority provides information on the disposal of hazardous waste.

Information on Packaging Materials

 PAP 22	 Plastics HDPE 2	 PP 5	 LDPE 4
---	--	---	--

11. SYMBOL KEY

	Manufacturer
	In vitro diagnostics
	Production lot number
	Usable until
	Temperature limit
	CE marking
	Article number
 Hinweis auf eIFU	Follow the Instruction For Use http://ifudownload.virotechdiagnostics.com

12. CHANGE HISTORY

<i>Revision</i>	<i>Section</i>	<i>Change</i>
<i>Rev 01</i>	<i>1-11</i>	<i>-</i>
<i>Rev 02</i>	<i>1-11</i>	<i>Editorial changes</i>
<i>Rev 03</i>	<i>1-12</i>	<i>Editorial changes, correction of disposal instructions, numbering of chapters added, change history created</i>